



Ref. Certif. No.

DE 3 - 3279

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST  
CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)  
CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE  
CERTIFICATS D'ESSAIS DES EQUIPEMENTS  
ELECTRIQUES (IECEE) METHODE OC

## CB TEST CERTIFICATE CERTIFICAT D'ESSAI OC

Product  
Produit

Equipment for measurement & control  
Fluid extraction system

Name and address of the applicant  
Nom et adresse du demandeur

Fluid Management Systems, Inc.  
580 Pleasant Street  
Watertown MA 02472, USA

Name and address of the manufacturer  
Nom et adresse du fabricant

Fluid Management Systems, Inc., 580 Pleasant Street, Watertown  
MA 02472, USA

Name and address of the factory  
Nom et adresse de l'usine

Fluid Management Systems, Inc., 580 Pleasant Street, Watertown  
MA 02472, USA

Rating and principal characteristics  
Valeurs nominales et caractéristiques principales

Rated Input Voltage: 120 V or 220 V  
Rated Frequency: 60 Hz or 50 Hz  
Rated Power: 350 or 200 W max (Control Module)  
1000 or 500 W max (Heater Block set)

Protection Class: I  
Degree of Protection: IPX0

Trade mark (if any)  
Marque de fabrique (si elle existe)

FMS

Model/type Ref.  
Ref. de type

PLE1, PLE2, PLE3  
PLE4, PLE5, PLE6

Additional information (if necessary)  
Information complémentaire (si nécessaire)

CBTL Procedure

A sample of the product was tested and found  
to be in conformity with  
Un échantillon de ce produit a été essayé et a été  
considéré conforme à la

IEC 61010-1:2001

as shown in the Test Report Ref. No.  
which form part of this certificate  
comme indiqué dans le Rapport d'essais numéro  
de référence qui constitue une partie de ce  
certificat

TÜV SÜD Product Service  
090-801252-000

This CB Test Certificate is issued by the National Certification Body  
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**

Date, 2008-07-29  
CB 08 07 67397 001

Joseph Janeliunas



TÜV SÜD Product Service GmbH · Certification Body · Ridlerstrasse 65 · D-80339 München

Product Service