



America

CERTIFICATE


No. U8 08 07 67397 002

Holder of Certificate: Fluid Management Systems, Inc.

580 Pleasant Street
Watertown MA 02472
USA

Production Facility(ies): 67397

Certification Mark:



Product: Equipment for measurement & control
Fluid extraction system

Model(s): PLE1, PLE2, PLE3
PLE4, PLE5, PLE6

Parameters:

Rated Input Voltage:	120 V or 220 V
Rated Frequency:	60 Hz or 50 Hz
Rated Power:	350 or 200 W max (Control Module) 1000 or 500 W max (Heater Block set)
Protection Class:	I
Degree of Protection:	IPX0

Tested according to: CAN/CSA C22.2 No. 61010-1:2004
UL 61010-1:2005
EN 61010-1:2001

The product was voluntarily tested according to the relevant safety requirements and mentioned properties. It can be marked with the certification mark shown above. The certification mark must not be altered in any way. See also notes overleaf.

Test report no.: 090-801252-000

Date, 2008-07-29

